

# G151 Datasheet



G·S·L



Non-sterile powder free synthetic gloves manufactured to conform to EN455 parts 1, 2 and 3 in compliance with Medical Devices Directive 93/42/EEC.

- Manufactured from synthetic rubber, latex free
- Minimum force at break (initial) 3.6N
- Tested to ASTM method, F1671-97
- Class 1 non-sterile CE classification
- AQL 1.5 freedom from holes
- Ambidextrous with beaded cuff for strength and ease of donning
- Available in four sizes

| Dimensions          | Small     | Medium    | Large     | Extra Large |
|---------------------|-----------|-----------|-----------|-------------|
| Palm Width (mm)     | 80+10mm   | 95+10mm   | 110+10mm  | >110mm      |
| Length (mm)         | Min 240mm | Min 240mm | Min 240mm | Min 240mm   |
| Thickness (mm) palm | 0.08mm    | 0.08mm    | 0.08mm    | 0.08mm      |
| Min force at break  | 3.6N      | 3.6N      | 3.6N      | 3.6N        |

| Benefits                |  |
|-------------------------|--|
| Extractable protein     | Free of the proteins found in natural rubber latex   |
| Accelerators            | Free of residual accelerators  |
| Viral penetration       | The gloves have been tested according to the ASTM method, F1671-97, and found to pass, thus proving that Gloveman Soft Touch is an effective barrier against penetration of viruses of size greater than 25-27nm |
| Quality standards       | Manufactured to conform to EN455 in compliance with Medical Devices Directive 93/42/EEC  |
| Sampling & inspection   | ISO2859  |
| Freedom from holes      | AQL 1.5  |
| CE classification Class | 1 non-sterile  |
| Shelf life              | 3 years  |

**Gloveman Supplies Ltd**  
 Vision House  
 Treleigh Ind Est  
 Redruth TR16 4AX  
 t. 01209 314759  
 f. 01209 314760  
 e. sales@gloveman.co.uk

[www.gloveman.co.uk](http://www.gloveman.co.uk)