

# G15 Datasheet



**G•S•L**

WWW.GLOVEMAN.CO.UK



Gloves tested to comply with Medical Devices Directives 93/42/EEC and European Standards EN455 parts 1, 2 and 3, EN388 and EN374. Designed for medical procedures, offering a high quality and dependable latex free glove.

- Visually inspected for holes and minor defects
- Random samples are filled with 1000ml of water and examined for leaks.
- Tested to AQL 1.5
- Manufactured from PVC plastisol suitable for medical and food contact applications
- Typical force at break: 4 newtons
- Typical elongation: 400%
- Typical tensile strength: 9 Mpa
- Available from size S to XL

Recommended for use with Medical, Laboratory, Chemical handling, Veterinary and Emergency services .

Dimensions	Small	Medium	Large	Extra Large
Palm Width (mm)	80+10mm	95+10mm	110+10mm	>110mm
Length (mm)	Min 240mm	Min 240mm	Min 240mm	Min 240mm
Thickness (mm) palm	0.09+0.02	0.09+0.02	0.09+0.02	0.09+0.02
Thickness (mm) finger	0.11+0.02	0.11+0.02	0.11+0.02	0.11+0.02
Thickness (mm) cuff	0.06+0.02	0.06+0.02	0.06+0.02	0.06+0.02

Benefits	
Latex free	Reducing the risk of hypersensitivity and allergic reactions
Soft, smooth formulation	For improved feel and fit
Powder free	Reduces potential dust contamination
Beaded cuff	For added strength and easier donning
Available in 4 sizes	Providing wearer with a range of comfortable sizes

Characteristics	Unaged	Aged
Force at break	4 Newtons	3.8 Newtons
Elongation at break	400%	380%

**Gloveman Supplies Ltd**  
 Vision House  
 Treleigh Ind Est  
 Redruth TR16 4AX  
 t. 01209 314759  
 f. 01209 314760  
 e. sales@gloveman.co.uk

[www.gloveman.co.uk](http://www.gloveman.co.uk)

Powder Free Vinyl Exam Gloves