

# G500 Datasheet Vynite Powder Free Gloves

AQL 1.5

EN455 Parts 1, 2, 3 & 4

ASTM Method F1671-97

Class 1 Medical Device

MDD 93/42/EEC



## General Details

Material	Nitrile and Polyvinyl Chloride
Surface Finish	Powder Free
Colour	Blue
Geometry	Ambidextrous
Texture	Smooth
Cuff and Finishing	Beaded

## Glove Specifications

Size	Weight g (M)	Length mm	Palm Width mm	Thickness		
				Cuff	Palm	Finger
XS	4 ± 0.3	230 (min)	75 ± 5	0.08	0.08	0.08
S	4.5 ± 0.3	230 (min)	85 ± 5	0.08	0.08	0.08
M	5 ± 0.3	230 (min)	95 ± 3	0.08	0.08	0.08
L	5.5 ± 0.3	230 (min)	105 ± 5	0.08	0.08	0.08
XL	6 ± 0.3	230 (min)	115 ± 5	0.08	0.08	0.08
Tensile strength at break			13MPa (min)			
Elongation at break			300%			

## Performance Requirements

Test	Related Defects	Inspection Level	AQL
Water Leak at Final Release	Holes	G-1 Single Sampling Plan	1.5

## Manufacturing Standards

Manufacturing Standards	Incompliance with ISO2859
Presentation	Standard: 100 gloves per dispenser box; 10 dispenser boxes per Carton
Glove Care	Should not store under direct sunlight or at high heat and humid conditions. Single use only. Avoid high polar solvents.



**G·S·L**

WWW.GLOVEMAN.CO.UK

**Gloveman Supplies Ltd**

Vision House, Treleigh Ind Est,

Redruth TR16 4AX

t. 01209 314759

f. 01209 314760